



GLOVES - NITRILE GLOVE, BOX OF 100

€18.00

SKU: 88500

Categories: [Bioide](#), [Gloves](#), [PersonProtect](#)

PRODUCT DESCRIPTION

Examination Gloves for single-use(Nitrile Gloves)

Sizes - S, M, L, XL

Cartoon Size - 240*125*63mm

EEC European food-grade full migration test

EN455-1 Test Report

ISO9001

Pictures are general, delivered products and packaging may look different.

[Technical Spec-nitrilr-glove](#)



SGS

Test Report

No. SHANGHAI13751401

Date: 01 Jul 2016

Page 1 of 3

The following sample(s) was/were submitted and identified on behalf of the clients as: Nitrile disposable glove

SGS Job No.: QNH.163000581CW-QD

Country of Origin: China

Date of Sample Received: 23 Jun 2016

Testing Period: 23 Jun 2016 - 01 Jul 2016

Test Requested: Selected test(s) as requested by client.

Test Method: Please refer to next page(s).

Test Results: Please refer to next page(s).

Result Summary:

Test Requested	Conclusion
Overall migration	PASS

Signed for and on behalf of
SGS-CTC Standards Technical Services (Shanghai) Co., Ltd.

Sanyu Wang
Approved Signature

Member of the SGS Group (2015-16)

SGS

Test Report

No. SHANGHAI13751401

Date: 01 Jul 2016

Page 2 of 3

Test Results:

Test Part Description:

Specimen No.	SGS Sample ID	Description	Material (Entered by the client)
SN1	SHAI16-137514-001	Blue nitrile RN	Butadiene

Overall migration

Test Requested: In accordance with Council of Europe Resolution AP (2004) 4 and EN 1186-2, to determine overall migration.

Test Method: With reference to Commission Regulation (EU) No 105/2011 of 14 January 2011 Annex II and Annex V for selection of condition and EN 1186-2:2002 for selection of test methods.

EN 1186-2:2002 value of 0.1 mg/dm² by total immersion method.

Detected Limit	Time	Temperature	Max. Permeation	Result of test	Conclusion
Residual value of 0.1 mg/dm²	2 (h/4h)	40°C	10 mg/dm²	< 0.1 mg/dm²	PASS

Notes:

(1) Analytical tolerance of test result is ± 0.1 mg/dm² or 20 mg/dm²

(2) Test condition & duration were specified by client.

SGS

Member of the SGS Group (2015-16)